



CERTIFICATE of Conformity

Reference No. : LCS170915019CS

Applicant : ZhuHai JiDa HuaPu Instrument Co.,Ltd

Address : 6/F, Build 5, #19, Yongtian Road, ZhuHai, Guangdong, China

Trademark : HoldPeak

Product : Multimeter

Model(s) : HP-870C, HP-870A, HP-870B, HP-870CR, HP-870D, HP-870E, HP-870F, HP-870G, HP-870H, HP-870F, HP-870M, HP-870MR, HP-870N, HP-570N, HP-570C-APP

Parameters : DC 9V for battery, CAT III 1000V

The submitted products have been tested by us with the listed standards and found in compliance with the following European Directives:

The LVD Directive 2014/35/EU

EN 61010-1: 2010

The tests were performed in normal operation mode. The test results apply only to the particular sample tested and to the specific tests carried out. This certificate applies specifically to the sample investigated in our test reference number only.

The CE markings as shown below can be affixed on the product after preparation of necessary technical documentation.

Other relevant Directives have to be observed.



Date of issue: November 03, 2017



Guangzhou LCS Compliance Testing Laboratory Ltd.
No 44-1, Qianfeng North Road, Shiqi town, Panyu District, Guangzhou City, China
Tel: (86)020-39166689 Fax: (86)020-39166619
Http://www.LCS-cert.com Email: webmaster@LCS-cert.com

