#### Soldering Hints

- (1) Put leads through mounting holes from the side with part outline. Ensue component evenly touch PCB.
- (2) Solder leads at the other side. Solder should fully fill and cover soldering pads. Avoid bridges between

Note:

□ R10

□ R11

□ R38

☐ R28, R40

R37, R39

Always meter resistor

values before soldering

□ R7, R36 : 180 Ω

□ R8, R12, R13: 120 Ω

□ R9. R15. R26: 1K Ω

□ L1,L3,L4 : 100 µ H

Cathode

□ D2

: 3K Ω

: 150 Ω

: 1.5K Ω

: 470 Ω

: 10K Ω

: 1N5819

: 1N4004 (or 1N4007)

neighbering pads. (3) Cut unused leads flush with cutter.

Resistors

ΠR1, R14, R16: 100K Ω

2. HF-Chokes

: 1.8M Ω

: 200K Ω

: 2M Ω

: 20K Ω

: 300 Ω


□ R2

□ R3

□ R4

□ R5

□ R6



## **DSO 138 Oscilloscope DIY Kit**

Applicable models: 13803K, 13804K User Manual Rev. 05 Applicable models: 13803K, 13804K Applicable firmware: 113-13801-060 or later

### Tools you need

- (1) Iron (20W) (4) Screw driver
- (2) Solder wire (5) Flush cutter

□ J4

Note:

optional.

SW6, SW7,

☐ C1, C9,

C10, C11,

C14, C15,

SW8

This connector is

☐ SW4, SW5, :6 X 6 X 5mm

(3) Multimeter (6) Tweezers

### Before you start -

- (1) Check part values & quantities against part list
- (2) Always meter resistor values before soldering
- (3) Understand all part polarities and orientations

# Important!!!

Install all SMD parts before proceeding to Step1 if you purchased kit 13804K.



# 15. Power connector

Solder positive pole

(the longer lead) to

the square pad



16. Pin-header (male) \*

17. Pin-header (female)

☐ J6

Note:

optional.

□ J7, J8

□ J3

14. Electrolytic capacitors

☐ C19, C21, : 100 µ F/16V

:DC005

:1 X 3 pin

:1 X 4 pin

: 1 X 2 pin

: 2 X 20 pin

These pin-headers are

C22, C24.

C25, C26

# 9. Pin header (for power)







## 10. Transistors



# Ceramic Capacitors

: 0.1 µ F

: 120pF

: USB mini -B



- □ U4
- :78L05 Attention!

# 11. Regulators



# 12. Capacitor trimmers



# 18. Slide switches



☐ SW1, SW2, : 2P3T

# 4. Crystal

3. Diodes





- C16, C17, C18, C20, C23
- □ C2 : 330pF □ C3 : 3pF
- ☐ C7, C8
  - C12, C13 : 22pF
- □ C5 : 1pF

Assembly Main Board and LCD board (follow the order as numbered)

5. USB Socket \*

6. Tact Switches



Solder positive pole (the longer lead) to □ D3 the square pad

:φ3mm, green

## 13. Power inductor



□ L2 : 1mH/0.5A

: 5 - 30pF

## 19. BNC connector



: BNC □ J1

### Note:

The thicker pins need to heat up longer to get good soldering result.

# 20. Test signal ring



- 1) Make a small ring with a lead cut-off.
- 2) Solder the ring to the two holes of J2 (as shown in the photo).





Note: Install to the side opposite to LCD panel.


☐ J1 :2 X 20 pin ☐ J2, J3 :1 X 2 pin

Notes: 1) JP1, JP2, JP5, and JP6 at bottom side should be kept open for normal running mode. 2) The USB connector do not have function. It was provided for future or user own use.


3) A 9V DC power supply (> 200mA capacity) is required to run the scope. Power supply is not included in the kit.


















### — How to Use -

### **Display and Controls**



### **Connections**

Power Supply: Connect DC power supply to J9 or J10. The power supply voltage must be in the range of 8 - 12V.

Probe: Connect probe to J1.

#### Attention

- Power supply voltage must not exceed
  12V. Otherwise U5 will get hot.
- Allowed maximum signal input voltage is 50Vpk (100Vpp) with the clip probe.

# **Operations**

Press on [SEL] button: Select parameter to be adjusted. The selected parameter will be highlighted.

Press on [+] or [-] button: Adjust the parameter selected by [SEL] button.

Press on [OK] button: Freeze waveform refresh (entering HOLD state). Press on it again will de-freeze.

Change [CPL] switch: Set couple to DC, AC, or GND. When GND is selected the scope input is isolated from input signal and connected to ground (0V input).

Change [SEN1] or [SEN2] switch: Adjust sensitivity. The product of [SEN1] and [SEN2] settings makes the actual sensitivity which is displayed at the lower-left corner of the panel.

Press on [Reset] button: Perform a system reset and re-boots the oscillscope.

#### Tips

#### Vpos Alignment

This is to fix the mismatch between 0V trace and VPos indicator. To do this set couple switch [CPL] to GND position. Press on [SEL] button to make VPos indicator highlighed. Hold down [OK] button for about 2 seconds. You will see VPos indicator aligned to 0V trace when you release [OK] button. You may see some residue mismatch remains at the highest sensitivity settings. This is normal.

Restore Factory Default

Hold down [+] and [-] buttons simultaneously for 2 seconds.

**Auto-center Trigger Level** 

Highlight trigger level indicator and hold down [OK] button for 2 seconds.


**Auto-center Horizontal Position** 


Highlight HPos indicator and hold down [OK] button for 2 seconds.


#### **Probe Calibration**

Because there is always some capacitance between scope input and ground probe needs to be calibrated to achieve better measurement results for high frequency signals. This can be done with the help of the built-in test signal. To do this please follow the steps below.

- 1. Connect the red clip to the test signal terminal and leave the black clip un-connected (see photo at right).
- 2. Set [SEN1] switch to 0.1V and [SEN2] switch to X5. Set [CPL] switch to AC or DC.
- Adjust timebase to 0.2ms. You should see waveform similar to that shown in photos below. If traces are not stable adjust trigger level (the pink triangle on right screen border) so as you get a stable display.
- 4. Turn C4 (capacitor trimmer) with a small screw driver so that the waveform displays sharp rightangle (photo C).
- Set [SEN1] switch to 1V and [SEN2] switch to X1while keep all other settings unchanged. Adjust C6 so that sharp rightangle waveform is displayed.







Leave black clip

Connect red clip to

test signal output

A - Not enough

B - Too much

C - Good

#### Turn On/Off Readouts

Press [SEL] so that timebase is highlighted. Hold down [OK] button for about 2 seconds. This will turn on/off measurement readouts.

#### Waveform Save/Recall

Press [SEL] & [+] simultaneously: Save currently displayed waveform to non-volatile memory. Press [SEL] & [-] simultaneously: Recall saved waveform

#### **Triggers and Their Modes**

Triggers are events that indicate signal voltage acrossing a set level (i.e. trigger level) along a specified direction (i.e. trigger slope, rising or falling). Oscilloscope uses triggers as reference points in time for stable waveform display and measurements.

#### Auto Mode

In auto mode oscilloscope will perform display refresh no matter triggers happen or not. When triggers are detected waveform display will be displayed with reference to trigger points. Otherwise, display waveform at ramdom reference points.

#### Normal Mode


In normal mode oscilloscope will only perform display refresh when there are triggers. If no triggers happen waveform display will stay unchanged.

#### Single Mode

Single mode is the same as normal mode except that oscilloscope will enter HOLD state after a trigger has been detected and waveform display has been updated.

Normal and single modes are useful for capturing sparse or single waveform.

Specifications	
Max realtime sample rate	1MSa/s
Analog bandwidth	0 200KHz
Sensitivity range	10mV/div - 5V/div
Max input voltage	50Vpk (1X probe)
Input impedance	1M ohm/20pF
Resolution	12 bits
Record length	1024 points
Timebase range	500s/Div 10us/Div
Trigger modes	Auto, Normal, and Single
Trigger position range	50%
Power supply	9V DC (8 – 12V)
Current consumption	~120mA
Dimension	117 x 76 x 15mm
Weight	70 gram (without probe)



Page 4